




America

CERTIFICATE

No. U8 069850 0011 Rev. 00

Holder of Certificate: Heidolph Instruments GmbH & Co. KG
Walpersdorfer Str. 12
91126 Schwabach
GERMANY

Certification Mark:



Product: Laboratory Equipment

Brand Name: heidolph

Model(s): Hei-VOLUME Distimatic Pro

Parameters:
Rated Voltage: 100 - 240 VAC
Rated Frequency: 50/60 Hz
Rated Power: max.1500 W
Protection Class: I

Tested according to:
UL 61010-1:2012/R:2018-11
CAN/CSA-C22.2 No. 61010-1:2012/A1:2018-11
UL 61010-2-010:2019
CSA-C22.2 No. 61010-2-010:2019
Supplemented by UL 61010-2-081:2019
Supplemented by CSA-C22.2 No. 61010-2-081:2019

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 713207810

Date, 2021-06-10

(Benedikt Pulver)